

発行日 2017年1月30日  
発行番号 17-01087-1

# 校正証明書

CALIBRATION CERTIFICATE

依頼者(Customer Name) 広島大学病院 総合医療研究推進センター 殿  
製品名(Description) おんどとり Jr. (株ティントデイ製)  
形式(Model No) RTR-501  
製造番号(Serial No) 52B80E70  
管理番号(Management No) 17809  
校正年月日(Date) 2017年1月26日

上記製品は、品質管理の国際規格ISO9001に適合した当社品質システムに基づき校正が行われ、校正作業における検査又は試験の結果は試験成績書通りです。  
この校正にかかわる当社使用標準器は、日本電気計器検定所をはじめとする公的校正機関の標準とトレーサビリティがとれております。

## 当社使用標準器

製品名 / 形式	製造番号	校正有効期限
デジタル指示温度計 / CAB-F200-8	0026/05	2017年10月31日
白金測温抵抗体 / NRHS1-0	TS6ZA 7100F	2017年10月31日

高山理化精機株式会社

〒399-0033 長野県松本市大字笹賀5652-18  
校正サービス課 TEL 0263-25-3111



発行番号

17-01087-1



# 試験成績書

依頼者 広島大学病院 総合医療研究推進センター 殿  
製品名 おんどとり Jr. (株式会社おんどとり製)  
形式 RTR-501  
製造番号 52B80E70  
管理番号 17809  
校正年月日 2017年1月26日  
校正時室温 24 °C 52 %

高山理化精機株式会社  
校正サービス課

上記製品の試験成績は次のとおりです。

校正結果

承認	校正者
	

試験温度(°C)

指示値(°C)

4.0

3.9

上記校正値は調整なしの結果です

使用標準器名称 / 形式	製造番号	精度
デジタル指示温度計 / CAB-F200-8	0026/05	±(0.15+0.002 t )°C
白金測温抵抗体 / NRHS1-0	TS6ZA 7100F	(精度は上記に含む)

上記使用標準器は、国家標準にトレーサブルであり又正常であることを確認済み。

以上

帳票登録番号 ; 409-01 Rev. A

Date 30/JAN/2017  
No 17-01087-9

# CALIBRATION CERTIFICATE



Customer Name Hiroshima University  
Description Thermo Recorder (T&D Corporation product)  
Model No RTR-501  
Serial No 52B80E70  
Management No 17809  
Date 26/JAN/2017

The above product was calibrated with the Takayama Rika Seiki Co., Ltd. quality system in accordance with the international standard of the quality control ISO9001, and the results of measurement or test in the calibration work are shown by the accompanying certificate of analysis.

The standards used in this calibration are traceability-ensured with standards of Japan Electric Meters Inspection Corporation (JEMIC) and other calibration organizations.

## Used Standards

Product Name / Type	Serial No	Duration of Calibration
Precision Thermometer / CAB-F200-8	0026/05	31/OCT/2017
Platinum Resistance Temperature Detector / NRHS1-0	TS6ZA 7100F	31/OCT/2017



  
**TAKAYAMA RIKA SEIKI CO., LTD.**  
5652-18, SASAGA, MATSUMOTO, NAGANO JAPAN 399-0033  
Calibration Service Department Tel: 263253111 

No 17-01087-9

# CERTIFICATE OF ANALYSIS

Customer Name Hiroshima University  
 Description Thermo Recorder (T&D Corporation product)  
 Model No RTR-501  
 Serial No 52B80E70  
 Management No 17809  
 Date 26/JAN/2017  
 Room Temperature  
     During Calibration 24 °C 52 %

  
 TAKAYAMA RIKI SEIKI CO., LTD.  
 Calibration Service Department

APPROVED	AUTHOR
	

Test data of the above product are as shown below;

### Calibration Result

Test Temperature (°C)	Reading (°C)
4.0	3.9

The above calibration data are with no adjustment.

Used STD Name / Type	Serial No	Accuracy
Precision Thermometer / CAB-F200-8	0026/05	± (0.15+0.002 t ) °C
Platinum Resistance Temperature Detector / NRHS1-0	TS6ZA 7100F	* Includes the accuracy of the Platinum Resistance Temperature Detector NRHS1-0

The above standards are traceability-ensured with national standards and are normalcy-confirmed.