


校正日：2020年2月28日
型式：RTR-502/TR-5106
製造番号：52BACDCF/-----

発行番号：S51916970
住友電気工業株式会社
横浜計測センター
横浜計測グループ



1. 機能点検

点検結果 良

2. 温度指示校正 (本体+センサ)

確度：±0.3℃ (-20~80℃)
±0.5℃ (-40~-20℃, 80~110℃)
±1.0℃ (-60~-40℃, 110~155℃)

試験点	基準範囲		指示値	判定
4℃	3.7℃	~ 4.3℃	4.3℃	良

Report Of Calibration

Reference No : S51916656

Sumitomo Electric Technical Solutions, Inc.
Yokohama Measurement Center
Yokohama Measurement Group

Customer's Name : Hiroshima University

Customer's Management No : -----

Description : Thermo Recorder
Manufacturer : T&D
Model No : RTR-502/TR-5106
Serial No : 52BACDCF/-----

Calibrated Date : 28-Feb-20
Temp : 23 °C
Status : PASSED
Adjusted : NONE

Inspector : R. Oobayashi
Rel. hum : 40 %
Control No : H071442

Approved by *J. Murahama*

This document certifies that the above instrument was calibrated in accordance with Japanese Industrial Standard, which is a STS operation standard.

The standard instrument we used for calibration is traceable to official calibration organizations such as Japan Electric Meters Inspection Corporation, Japan Quality Assurance Organization, and official calibration organizations overseas which joined the national Metrological Committee, such as the National Institute of Standards and Technology.

【Calibration Equipment Used】

Description	Manufacturer	Trace No
Model	Serial No	Expiration Date
Digital Thermometer	HART SCIENTIFIC	5K050090
1502A/5615	A15441/833504	31-Jul-20

Calibrated Date : 28-Feb-20
Model No : RTR-502/TR-5106
Serial No : 52BACDCF/-----

Reference No : S51916656
Sumitomo Electric Technical Solutions, Inc.
Yokohama Measurement Center
Yokohama Measurement Group

1. Functional Check Results Pass

2. Temperature Instructions Test (Base+Sensor)

Accuracy : $\pm 0.3^{\circ}\text{C}$ (-20~80°C)
 $\pm 0.5^{\circ}\text{C}$ (-40~-20°C, 80~110°C)
 $\pm 1.0^{\circ}\text{C}$ (-60~-40°C, 110~155°C)

Test point	Limits		Indicated value	Status
4 °C	3.7 °C ~	4.3 °C	4.3 °C	Pass